

SUPERVISOR TEST

WASHINGTON CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 100284

Date: 08/29/2022

OPERATOR NAME:
MICHAEL O WEBER
Permit #: 46067
Operator Agency/Dept: WASHINGTON CO. SO
Reference Standard #: 22190

Nominal = 0.10 g/210L

BLANK TEST	0.00	06:46
INTERNAL STANDARD	VERIFIED	06:46
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	06:47
BLANK TEST	0.00	06:48
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	06:48
BLANK TEST	0.00	06:49
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	06:49
BLANK TEST	0.00	06:50

