

SUPERVISOR TEST

WASHINGTON CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 100284

Date: 08/15/2022

OPERATOR NAME:
MICHAEL O WEBER
Permit #: 46067
Operator Agency/Dept: WASHINGTON CO. SO
Reference Standard #: 22190

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:08
INTERNAL STANDARD	VERIFIED	07:08
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	07:08
BLANK TEST	0.00	07:09
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	07:09
BLANK TEST	0.00	07:10
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	07:10
BLANK TEST	0.00	07:11

