

SUPERVISOR TEST

HUDSON FALLS
 VILLAGE POLICE DEPT.
 DATAMASTER dmt: 102408

Date: 09/07/2022

OPERATOR NAME:
 DAVID M COSTELLO
 Permit #: 40891
 Operator Agency/Dept: HUDSON FALLS VIL. PD
 Reference Standard #: 22190

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:28
INTERNAL STANDARD	VERIFIED	03:28
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	03:29
BLANK TEST	0.00	03:30
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	03:30
BLANK TEST	0.00	03:31
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	03:31
BLANK TEST	0.00	03:32

