

SOLUTION CHANGE

 HUDSON FALLS
 VILLAGE POLICE DEPT.
 DATAMASTER dmt: 102408

Date: 09/12/2022

OPERATOR NAME:
 DAVID M COSTELLO
 Permit #: 40891
 Operator Agency/Dept: HUDSON FALLS VIL. PD
 Reference Standard #: 22190

Nominal = 0.10 g/210L

BLANK TEST	0.00	19:14
INTERNAL STANDARD	VERIFIED	19:14
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	19:14
BLANK TEST	0.00	19:15
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	19:16
BLANK TEST	0.00	19:16
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	19:17
BLANK TEST	0.00	19:18

