

SUPERVISOR TEST

GRANVILLE VILLAGE
POLICE DEPT
DATAMASTER dmt: 113306

Date: 09/12/2022

OPERATOR NAME:
JASON M VANDENBURGH
Permit #: 37635
Operator Agency/Dept: GRANVILLE VILLAGE PD
Reference Standard #: 22190

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:16
INTERNAL STANDARD	VERIFIED	08:16
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	08:16
BLANK TEST	0.00	08:17
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	08:18
BLANK TEST	0.00	08:18
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	08:19
BLANK TEST	0.00	08:20

