

SUPERVISOR TEST

GRANVILLE VILLAGE
POLICE DEPT
DATAMASTER dmt: 113306

Date: 09/19/2022

OPERATOR NAME:
ERNEST R BASSETT
Permit #: 25518
Operator Agency/Dept: GRANVILLE VILLAGE PD
Reference Standard #: 22190

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:17
INTERNAL STANDARD	VERIFIED	08:17
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	08:17
BLANK TEST	0.00	08:18
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	08:18
BLANK TEST	0.00	08:19
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	08:20
BLANK TEST	0.00	08:20

