

SUPERVISOR TEST

WHITEHALL VILLAGE
 POLICE DEPT.
 DATAMASTER dmt: 113806

Date: 09/17/2022

OPERATOR NAME:
 BRYAN J GRECO
 Permit #: 44077
 Operator Agency/Dept: WHITEHALL VILL. P.D.
 Reference Standard #: 22190

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:21
INTERNAL STANDARD	VERIFIED	01:22
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	01:22
BLANK TEST	0.00	01:23
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	01:23
BLANK TEST	0.00	01:24
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	01:24
BLANK TEST	0.00	01:25

