

SOLUTION CHANGE

FORT EDWARD
VILLAGE POLICE DEPT.
DATAMASTER dmt: 115708

Date: 09/05/2022

OPERATOR NAME:
JOSEPH P DEANGELO
Permit #: 49218
Operator Agency/Dept: FORT EDWARD VILL. PD
Reference Standard #: 22190

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:14
INTERNAL STANDARD	VERIFIED	23:14
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	23:14
BLANK TEST	0.00	23:15
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	23:16
BLANK TEST	0.00	23:17
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	23:17
BLANK TEST	0.00	23:18

JD
#1203

