

SUPERVISOR TEST

FORT EDWARD
VILLAGE POLICE DEPT.
DATAMASTER dmt: 115708

Date: 09/12/2022

OPERATOR NAME:
JOSEPH P DEANGELO
Permit #: 49218
Operator Agency/Dept: FORT EDWARD VILL. PD
Reference Standard #: 22190

Nominal = 0.10 g/210L

BLANK TEST	0.00	13:42
INTERNAL STANDARD	VERIFIED	13:42
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	13:43
BLANK TEST	0.00	13:44
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	13:44
BLANK TEST	0.00	13:45
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	13:45
BLANK TEST	0.00	13:46

JD
#1203

