

SOLUTION CHANGE

 WASHINGTON CO.
 SHERIFF'S OFFICE
 DATAMASTER dmt: 100284

Date: 09/03/2022

OPERATOR NAME:
 ROBERT J SULLIVAN
 Permit #: 34106
 Operator Agency/Dept: WASHINGTON CO. SO
 Reference Standard #: 22190

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:43
INTERNAL STANDARD	VERIFIED	08:43
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	08:43
BLANK TEST	0.00	08:44
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	08:44
BLANK TEST	0.00	08:45
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	08:45
BLANK TEST	0.00	08:46

