

SUPERVISOR TEST

WASHINGTON CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 100284

Date: 09/07/2022

OPERATOR NAME:
ROBERT J SULLIVAN
Permit #: 34106
Operator Agency/Dept: WASHINGTON CO. SO
Reference Standard #: 22190

Nominal = 0.10 g/210L

BLANK TEST	0.00	06:49
INTERNAL STANDARD	VERIFIED	06:50
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	06:50
BLANK TEST	0.00	06:51
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	06:51
BLANK TEST	0.00	06:52
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	06:52
BLANK TEST	0.00	06:53

