

SUPERVISOR TEST

GRANVILLE VILLAGE
POLICE DEPT
DATAMASTER dmt: 113306

Date: 09/26/2022

OPERATOR NAME:
ERNEST R BASSETT
Permit #: 25518
Operator Agency/Dept: GRANVILLE VILLAGE PD
Reference Standard #: 22190

Nominal = 0.10 g/210L

BLANK TEST	0.00	06:27
INTERNAL STANDARD	VERIFIED	06:27
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	06:28
BLANK TEST	0.00	06:29
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	06:29
BLANK TEST	0.00	06:30
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	06:30
BLANK TEST	0.00	06:31

