

SOLUTION CHANGE

-----  
 WHITEHALL VILLAGE  
 POLICE DEPT.  
 DATAMASTER dmt: 113806  
 -----

Date: 10/02/2022

OPERATOR NAME:  
 BRYAN J GRECO  
 Permit #: 44077  
 Operator Agency/Dept: WHITEHALL VILL. P.D.  
 Reference Standard #: 22260

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:19
INTERNAL STANDARD	VERIFIED	01:19
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	01:19
BLANK TEST	0.00	01:20
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	01:20
BLANK TEST	0.00	01:21
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	01:21
BLANK TEST	0.00	01:22

