

SUPERVISOR TEST

-----  
WASHINGTON COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 106406  
-----

Date: 09/27/2022

OPERATOR NAME:  
ROBERT J SULLIVAN  
Permit #: 34106  
Operator Agency/Dept: WASHINGTON CO. SO  
Reference Standard #: 22190

Nominal = 0.10 g/210L

BLANK TEST	0.00	12:24
INTERNAL STANDARD	VERIFIED	12:24
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	12:24
BLANK TEST	0.00	12:25
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	12:25
BLANK TEST	0.00	12:26
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	12:27
BLANK TEST	0.00	12:28

