

SUPERVISOR TEST

WASHINGTON CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 100284

Date: 09/26/2022

OPERATOR NAME:
MICHAEL O WEBER
Permit #: 46067
Operator Agency/Dept: WASHINGTON CO. SO
Reference Standard #: 22190

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:45
INTERNAL STANDARD	VERIFIED	07:45
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	07:46
BLANK TEST	0.00	07:46
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	07:47
BLANK TEST	0.00	07:48
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	07:48
BLANK TEST	0.00	07:49

