

SUPERVISOR TEST

WHITEHALL VILLAGE  
POLICE DEPT.  
DATAMASTER dmt: 113806

Date: 10/08/2022

OPERATOR NAME:  
ALEXANDER T GILMORE  
Permit #: 47649  
Operator Agency/Dept: WHITEHALL VILL. P.D.  
Reference Standard #: 22260

Nominal = 0.10 g/210L

BLANK TEST	0.00	14:05
INTERNAL STANDARD	VERIFIED	14:05
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	14:06
BLANK TEST	0.00	14:07
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	14:07
BLANK TEST	0.00	14:08
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	14:08
BLANK TEST	0.00	14:09

