

SUPERVISOR TEST

GRANVILLE VILLAGE
POLICE DEPT
DATAMASTER dmt: 113306

Date: 10/17/2022

OPERATOR NAME:
ERNEST R BASSETT
Permit #: 25518
Operator Agency/Dept: GRANVILLE VILLAGE PD
Reference Standard #: 22260

Nominal = 0.10 g/210L

BLANK TEST	0.00	06:15
INTERNAL STANDARD	VERIFIED	06:15
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	06:15
BLANK TEST	0.00	06:16
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	06:16
BLANK TEST	0.00	06:17
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	06:18
BLANK TEST	0.00	06:19

