

SUPERVISOR TEST

WASHINGTON CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 100284

Date: 10/17/2022

OPERATOR NAME:  
MICHAEL O WEBER  
Permit #: 46067  
Operator Agency/Dept: WASHINGTON CO. SO  
Reference Standard #: 22260

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:17
INTERNAL STANDARD	VERIFIED	07:17
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	07:18
BLANK TEST	0.00	07:18
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	07:19
BLANK TEST	0.00	07:19
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	07:20
BLANK TEST	0.00	07:21

