

SOLUTION CHANGE

 WASHINGTON COUNTY
 SHERIFF'S OFFICE
 DATAMASTER dmt: 106406

Date: 10/03/2022

OPERATOR NAME:
 CORBET W SULLIVAN
 Permit #: 45898
 Operator Agency/Dept: WASHINGTON CO. SO
 Reference Standard #: 22260

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:13
INTERNAL STANDARD	VERIFIED	07:13
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	07:14
BLANK TEST	0.00	07:15
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	07:15
BLANK TEST	0.00	07:16
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	07:16
BLANK TEST	0.00	07:17

