

SUPERVISOR TEST

WASHINGTON COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 106406

Date: 10/11/2022

OPERATOR NAME:
KYLE L KOLAR
Permit #: 44063
Operator Agency/Dept: WASHINGTON CO. SO
Reference Standard #: 22260

Nominal = 0.10 g/210L

BLANK TEST	0.00	20:11
INTERNAL STANDARD	VERIFIED	20:12
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	20:12
BLANK TEST	0.00	20:13
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	20:13
BLANK TEST	0.00	20:14
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	20:14
BLANK TEST	0.00	20:15

