

SUPERVISOR TEST

GRANVILLE VILLAGE
POLICE DEPT
DATAMASTER dmt: 113306

Date: 10/22/2022

OPERATOR NAME:
ERNEST R BASSETT
Permit #: 25518
Operator Agency/Dept: GRANVILLE VILLAGE PD
Reference Standard #: 22260

Nominal = 0.10 g/210L

BLANK TEST	0.00	06:25
INTERNAL STANDARD	VERIFIED	06:25
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	06:26
BLANK TEST	0.00	06:27
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	06:27
BLANK TEST	0.00	06:28
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	06:28
BLANK TEST	0.00	06:29

