

SUPERVISOR TEST

WHITEHALL VILLAGE  
 POLICE DEPT.  
 DATAMASTER dmt: 113806

Date: 10/22/2022

OPERATOR NAME:  
 ALEXANDER T GILMORE  
 Permit #: 47649  
 Operator Agency/Dept: WHITEHALL VILL. P.D.  
 Reference Standard #: 22260

Nominal = 0.10 g/210L

|                         |          |       |
|-------------------------|----------|-------|
| BLANK TEST              | 0.00     | 13:21 |
| INTERNAL STANDARD       | VERIFIED | 13:21 |
| REFERENCE STANDARD TEMP | 33.99c   |       |
| REFERENCE STANDARD      | 0.10     | 13:22 |
| BLANK TEST              | 0.00     | 13:23 |
| REFERENCE STANDARD TEMP | 33.99c   |       |
| REFERENCE STANDARD      | 0.10     | 13:23 |
| BLANK TEST              | 0.00     | 13:24 |
| REFERENCE STANDARD TEMP | 33.99c   |       |
| REFERENCE STANDARD      | 0.10     | 13:24 |
| BLANK TEST              | 0.00     | 13:25 |

