

SOLUTION CHANGE

GRANVILLE VILLAGE
POLICE DEPT
DATAMASTER dmt: 113306

Date: 11/07/2022

OPERATOR NAME:
JASON M VANDENBURGH
Permit #: 37635
Operator Agency/Dept: GRANVILLE VILLAGE PD
Reference Standard #: 22260

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:44
INTERNAL STANDARD	VERIFIED	07:44
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	07:44
BLANK TEST	0.00	07:45
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	07:46
BLANK TEST	0.00	07:47
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	07:47
BLANK TEST	0.00	07:48

