

SOLUTION CHANGE

WHITEHALL VILLAGE
 POLICE DEPT.
 DATAMASTER dmt: 113806

Date: 11/05/2022

OPERATOR NAME:
 BRYAN J GRECO
 Permit #: 44077
 Operator Agency/Dept: WHITEHALL VILL. P.D.
 Reference Standard #: 22260

Nominal = 0.10 g/210L

BLANK TEST	0.00	12:11
INTERNAL STANDARD	VERIFIED	12:11
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	12:11
BLANK TEST	0.00	12:12
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	12:13
BLANK TEST	0.00	12:14
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	12:14
BLANK TEST	0.00	12:15

