

SUPERVISOR TEST

WHITEHALL VILLAGE
 POLICE DEPT.
 DATAMASTER dmt: 113806

Date: 11/19/2022

OPERATOR NAME:
 ALEXANDER T GILMORE
 Permit #: 47649
 Operator Agency/Dept: WHITEHALL VILL. P.D.
 Reference Standard #: 22260

Nominal = 0.10 g/210L

BLANK TEST	0.00	13:58
INTERNAL STANDARD	VERIFIED	13:58
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	13:59
BLANK TEST	0.00	13:59
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	14:00
BLANK TEST	0.00	14:01
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	14:01
BLANK TEST	0.00	14:02

