

SOLUTION CHANGE

-----  
CAMBRIDGE VILLAGE  
POLICE DEPT  
DATAMASTER dmt: 126706  
-----

Date: 11/01/2022

OPERATOR NAME:  
ROBERT M DANKO  
Permit #: 39105  
Operator Agency/Dept: CAMBRIDGE VILL. PD  
Reference Standard #: 22260

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:56
INTERNAL STANDARD	VERIFIED	08:56
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	08:57
BLANK TEST	0.00	08:57
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	08:58
BLANK TEST	0.00	08:58
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	08:59
BLANK TEST	0.00	09:00

