

SUPERVISOR TEST

 FORT EDWARD
 VILLAGE POLICE DEPT.
 DATAMASTER dmt: 115708

Date: 10/31/2022

OPERATOR NAME:
 JOSEPH P DEANGELO
 Permit #: 49218
 Operator Agency/Dept: FORT EDWARD VILL. PD
 Reference Standard #: 22260

Nominal = 0.10 g/210L

BLANK TEST	0.00	10:46
INTERNAL STANDARD	VERIFIED	10:46
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	10:46
BLANK TEST	0.00	10:47
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	10:48
BLANK TEST	0.00	10:48
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	10:49
BLANK TEST	0.00	10:50

JD
 #1283

