

SUPERVISOR TEST

-----  
FORT EDWARD  
VILLAGE POLICE DEPT.  
DATAMASTER dmt: 115708  
-----

Date: 10/19/2022

OPERATOR NAME:  
JOSEPH P DEANGELO  
Permit #: 49218  
Operator Agency/Dept: FORT EDWARD VILL. PD  
Reference Standard #: 22260

Nominal = 0.10 g/210L

BLANK TEST	0.00	18:32
INTERNAL STANDARD	VERIFIED	18:33
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	18:33
BLANK TEST	0.00	18:34
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	18:34
BLANK TEST	0.00	18:35
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	18:36
BLANK TEST	0.00	18:36

*JD*  
*#1203*

