

SUPERVISOR TEST

FORT EDWARD
VILLAGE POLICE DEPT.
DATAMASTER dmt: 115708

Date: 10/25/2022

OPERATOR NAME:
JOSEPH P DEANGELO
Permit #: 49218
Operator Agency/Dept: FORT EDWARD VILL. PD
Reference Standard #: 22260

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:41
INTERNAL STANDARD	VERIFIED	22:41
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	22:41
BLANK TEST	0.00	22:42
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	22:42
BLANK TEST	0.00	22:43
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	22:44
BLANK TEST	0.00	22:45

JD
#1203

