

SUPERVISOR TEST

GRANVILLE VILLAGE  
POLICE DEPT  
DATAMASTER dmt: 113306

Date: 11/28/2022

OPERATOR NAME:  
ERNEST R BASSETT  
Permit #: 25518  
Operator Agency/Dept: GRANVILLE VILLAGE PD  
Reference Standard #: 22260

Nominal = 0.10 g/210L

BLANK TEST	0.00	05:42
INTERNAL STANDARD	VERIFIED	05:42
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	05:43
BLANK TEST	0.00	05:44
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	05:44
BLANK TEST	0.00	05:45
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	05:45
BLANK TEST	0.00	05:46

