

SUPERVISOR TEST

WASHINGTON COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 106406

Date: 11/28/2022

OPERATOR NAME:
MATTHEW K JACKSON
Permit #: 40078
Operator Agency/Dept: WASHINGTON CO. SO
Reference Standard #: 22260

Nominal = 0.10 g/210L

BLANK TEST	0.00	15:45
INTERNAL STANDARD	VERIFIED	15:45
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	15:45
BLANK TEST	0.00	15:46
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	15:46
BLANK TEST	0.00	15:47
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	15:48
BLANK TEST	0.00	15:48

