

SOLUTION CHANGE

WASHINGTON COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 106406

Date: 11/03/2022

OPERATOR NAME:
ROBERT J SULLIVAN
Permit #: 34106
Operator Agency/Dept: WASHINGTON CO. SO
Reference Standard #: 22260

Nominal = 0.10 g/210L

BLANK TEST	0.00	15:06
INTERNAL STANDARD	VERIFIED	15:06
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	15:06
BLANK TEST	0.00	15:07
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	15:08
BLANK TEST	0.00	15:08
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	15:09
BLANK TEST	0.00	15:10

