

SUPERVISOR TEST

WASHINGTON COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 106406

Date: 11/14/2022

OPERATOR NAME:  
PAUL S HENRY  
Permit #: 49918  
Operator Agency/Dept: WASHINGTON CO. SO  
Reference Standard #: 22260

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:46
INTERNAL STANDARD	VERIFIED	23:46
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	23:46
BLANK TEST	0.00	23:47
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	23:47
BLANK TEST	0.00	23:48
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	23:49
BLANK TEST	0.00	23:49

*Paul Henry #640*

