

SUPERVISOR TEST

WASHINGTON CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 100284

Date: 11/29/2022

OPERATOR NAME:  
ERIC C WRIGHT  
Permit #: 49047  
Operator Agency/Dept: WASHINGTON CO. SO  
Reference Standard #: 22260

Nominal = 0.10 g/210L

BLANK TEST	0.00	06:35
INTERNAL STANDARD	VERIFIED	06:35
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	06:35
BLANK TEST	0.00	06:36
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	06:37
BLANK TEST	0.00	06:37
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	06:38
BLANK TEST	0.00	06:39

