

SUPERVISOR TEST

WASHINGTON CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 100284

Date: 10/24/2022

OPERATOR NAME:
CARSON D MCKINLEY
Permit #: 51169
Operator Agency/Dept: WASHINGTON CO. SO
Reference Standard #: 22260

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:05
INTERNAL STANDARD	VERIFIED	07:05
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	07:05
BLANK TEST	0.00	07:06
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	07:07
BLANK TEST	0.00	07:07
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	07:08
BLANK TEST	0.00	07:08

