

SUPERVISOR TEST

WASHINGTON CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 100284

Date: 11/07/2022

OPERATOR NAME:
MICHAEL O WEBER
Permit #: 46067
Operator Agency/Dept: WASHINGTON CO. SO
Reference Standard #: 22260

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:08
INTERNAL STANDARD	VERIFIED	08:08
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	08:08
BLANK TEST	0.00	08:09
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	08:09
BLANK TEST	0.00	08:10
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	08:10
BLANK TEST	0.00	08:11

