

SUPERVISOR TEST

WASHINGTON CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 100284

Date: 11/14/2022

OPERATOR NAME:
TREVOR J PECK
Permit #: 49747
Operator Agency/Dept: WASHINGTON CO. SO
Reference Standard #: 22260

Nominal = 0.10 g/210L

BLANK TEST	0.00	05:53
INTERNAL STANDARD	VERIFIED	05:53
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	05:54
BLANK TEST	0.00	05:55
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	05:55
BLANK TEST	0.00	05:56
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	05:56
BLANK TEST	0.00	05:57

11/14/2022

#637

