

SUPERVISOR TEST

 FORT EDWARD
 VILLAGE POLICE DEPT.
 DATAMASTER dmt: 115708

Date: 11/28/2022

OPERATOR NAME:
 JOSEPH P DEANGELO
 Permit #: 49218
 Operator Agency/Dept: FORT EDWARD VILL. PD
 Reference Standard #: 22260

Nominal = 0.10 g/210L

BLANK TEST	0.00	15:24
INTERNAL STANDARD	VERIFIED	15:24
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	15:24
BLANK TEST	0.00	15:25
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	15:26
BLANK TEST	0.00	15:27
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	15:27
BLANK TEST	0.00	15:28

JD
 #1203

