

SUPERVISOR TEST

FORT EDWARD
VILLAGE POLICE DEPT.
DATAMASTER dmt: 115708

Date: 11/14/2022

OPERATOR NAME:
JOSEPH P DEANGELO
Permit #: 49218
Operator Agency/Dept: FORT EDWARD VILL. PD
Reference Standard #: 22260

Nominal = 0.10 g/210L

BLANK TEST	0.00	15:23
INTERNAL STANDARD	VERIFIED	15:23
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	15:24
BLANK TEST	0.00	15:25
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	15:25
BLANK TEST	0.00	15:26
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	15:26
BLANK TEST	0.00	15:27

JD
#1203

