

SUPERVISOR TEST

-----  
CAMBRIDGE VILLAGE  
POLICE DEPT  
DATAMASTER dmt: 126706  
-----

Date: 11/28/2022

OPERATOR NAME:  
ROBERT M DANKO  
Permit #: 39105  
Operator Agency/Dept: CAMBRIDGE VILL. PD  
Reference Standard #: 22260

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:03
INTERNAL STANDARD	VERIFIED	08:03
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	08:04
BLANK TEST	0.00	08:05
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	08:05
BLANK TEST	0.00	08:06
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	08:06
BLANK TEST	0.00	08:07

