

SUPERVISOR TEST

CAMBRIDGE VILLAGE
 POLICE DEPT
 DATAMASTER dmt: 126706

Date: 11/14/2022

OPERATOR NAME:
 ROBERT M DANKO
 Permit #: 39105
 Operator Agency/Dept: CAMBRIDGE VILL. PD
 Reference Standard #: 22260

Nominal = 0.10 g/210L

BLANK TEST	0.00	06:51
INTERNAL STANDARD	VERIFIED	06:51
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	06:51
BLANK TEST	0.00	06:52
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	06:52
BLANK TEST	0.00	06:53
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	06:54
BLANK TEST	0.00	06:54

