

SOLUTION CHANGE

 WHITEHALL VILLAGE
 POLICE DEPT.
 DATAMASTER dmt: 113806

Date: 12/03/2022

OPERATOR NAME:
 BRYAN J GRECO
 Permit #: 44077
 Operator Agency/Dept: WHITEHALL VILL. P.D.
 Reference Standard #: 22390

Nominal = 0.10 g/210L

BLANK TEST	0.00	14:51
INTERNAL STANDARD	VERIFIED	14:52
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	14:52
BLANK TEST	0.00	14:53
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	14:53
BLANK TEST	0.00	14:54
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	14:54
BLANK TEST	0.00	14:55

