

SUPERVISOR TEST

GRANVILLE VILLAGE
POLICE DEPT
DATAMASTER dmt: 113306

Date: 12/05/2022

OPERATOR NAME:
JASON M VANDENBURGH
Permit #: 37635
Operator Agency/Dept: GRANVILLE VILLAGE PD
Reference Standard #: 22260

Nominal = 0.10 g/210L

BLANK TEST	0.00	10:11
INTERNAL STANDARD	VERIFIED	10:11
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	10:11
BLANK TEST	0.00	10:12
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	10:12
BLANK TEST	0.00	10:13
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	10:14
BLANK TEST	0.00	10:14

