

SUPERVISOR TEST

GRANVILLE VILLAGE
POLICE DEPT
DATAMASTER dmt: 113306

Date: 12/19/2022

OPERATOR NAME:
JASON M VANDENBURGH
Permit #: 37635
Operator Agency/Dept: GRANVILLE VILLAGE PD
Reference Standard #: 22260

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:20
INTERNAL STANDARD	VERIFIED	07:20
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	07:20
BLANK TEST	0.00	07:21
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	07:21
BLANK TEST	0.00	07:22
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	07:23
BLANK TEST	0.00	07:23

