

SUPERVISOR TEST

 WHITEHALL VILLAGE
 POLICE DEPT.
 DATAMASTER dmt: 113806

Date: 12/24/2022

OPERATOR NAME:
 LOGAN P MARINO
 Permit #: 50153
 Operator Agency/Dept: WHITEHALL VILL. P.D.
 Reference Standard #: 22390

Nominal = 0.10 g/210L

BLANK TEST	0.00	06:39
INTERNAL STANDARD	VERIFIED	06:39
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	06:40
BLANK TEST	0.00	06:41
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	06:41
BLANK TEST	0.00	06:42
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	06:42
BLANK TEST	0.00	06:43

