

SUPERVISOR TEST

WHITEHALL VILLAGE
POLICE DEPT.
DATAMASTER dmt: 113806

Date: 12/31/2022

OPERATOR NAME:
BRYAN J GRECO
Permit #: 44077
Operator Agency/Dept: WHITEHALL VILL. P.D.
Reference Standard #: 22390

Nominal = 0.10 g/210L

BLANK TEST	0.00	06:00
INTERNAL STANDARD	VERIFIED	06:00
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	06:00
BLANK TEST	0.00	06:01
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	06:02
BLANK TEST	0.00	06:03
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	06:03
BLANK TEST	0.00	06:04

