

SOLUTION CHANGE

WASHINGTON COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 106406

Date: 01/02/2023

OPERATOR NAME:
LOGAN T GILLIS
Permit #: 51159
Operator Agency/Dept: WASHINGTON CO. SO
Reference Standard #: 22390

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:23
INTERNAL STANDARD	VERIFIED	23:24
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	23:24
BLANK TEST	0.00	23:25
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	23:25
BLANK TEST	0.00	23:26
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	23:26
BLANK TEST	0.00	23:27

LG#632

