

SOLUTION CHANGE

 WASHINGTON CO.
 SHERIFF'S OFFICE
 DATAMASTER dmt: 100284

Date: 01/02/2023

OPERATOR NAME:
 ROBERT J SULLIVAN
 Permit #: 34106
 Operator Agency/Dept: WASHINGTON CO. SO
 Reference Standard #: 22390

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:14
INTERNAL STANDARD	VERIFIED	09:14
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	09:14
BLANK TEST	0.00	09:15
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	09:15
BLANK TEST	0.00	09:16
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	09:16
BLANK TEST	0.00	09:17

