

SOLUTION CHANGE

CAMBRIDGE VILLAGE  
POLICE DEPT  
DATAMASTER dmt: 126706

Date: 10/03/2022

OPERATOR NAME:  
ROBERT M DANKO  
Permit #: 39105  
Operator Agency/Dept: CAMBRIDGE VILL. PD  
Reference Standard #: 22190

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:00
INTERNAL STANDARD	VERIFIED	09:00
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	09:01
BLANK TEST	0.00	09:02
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	09:02
BLANK TEST	0.00	09:03
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	09:03
BLANK TEST	0.00	09:04

