

SUPERVISOR TEST

-----  
 CAMBRIDGE VILLAGE  
 POLICE DEPT  
 DATAMASTER dmt: 126706  
 -----

Date: 01/09/2023

OPERATOR NAME:  
 ROBERT M DANKO  
 Permit #: 39105  
 Operator Agency/Dept: CAMBRIDGE VILL. PD  
 Reference Standard #: 22390

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:11
INTERNAL STANDARD	VERIFIED	09:11
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	09:12
BLANK TEST	0.00	09:12
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	09:13
BLANK TEST	0.00	09:14
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	09:14
BLANK TEST	0.00	09:15

